

***E. coli* (Protein Expression Strains) HCP ELISA Kit
(One-step ELISA)**Product No.: **1301301-1**Lot No.: **20250803**Date of Mfg.: **26 Aug. 2025**Expiration Date: **15 Jul. 2026**

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Limit of Quantification	1.5 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration was not more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.200.	PASS

Approved By Cheng JiajunDAY/MONTH/YEAR 27/08/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China